

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050725\
 Data File : VX046075.D
 Acq On : 07 May 2025 16:13
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: May 08 01:49:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	68216	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	137093	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	129625	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	55978	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67528	53.098	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.200%
35) Dibromofluoromethane	5.385	113	49641	50.284	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.560%
50) Toluene-d8	8.647	98	172164	50.386	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.780%
62) 4-Bromofluorobenzene	11.079	95	67929	51.828	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050725\
Data File : VX046075.D
Acq On : 07 May 2025 16:13
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: May 08 01:49:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

