

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050819\
 Data File : VX009397.D
 Acq On : 08 May 2019 14:46
 Operator : JC/SP
 Sample : K2736-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-190507

Quant Time: May 09 06:53:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 03:31:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	179389	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	286021	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	247360	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	103808	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	117510	47.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.98%	
35) Dibromofluoromethane	5.50	113	87550	48.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.38%	
50) Toluene-d8	8.71	98	354809	48.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.76%	
62) 4-Bromofluorobenzene	11.13	95	114923	43.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.14%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	17706	8.126	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050819\
Data File : VX009397.D
Acq On : 08 May 2019 14:46
Operator : JC/SP
Sample : K2736-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-190507

Quant Time: May 09 06:53:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 03:31:13 2019
Response via : Initial Calibration

