

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050819\
 Data File : VX009399.D
 Acq On : 08 May 2019 15:33
 Operator : JC/SP
 Sample : PB119521TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB119521TB

Quant Time: May 09 06:56:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 03:31:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	181705	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	293473	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	259476	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	112945	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	120377	48.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.06%	
35) Dibromofluoromethane	5.50	113	89391	48.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.90%	
50) Toluene-d8	8.71	98	372682	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
62) 4-Bromofluorobenzene	11.13	95	122526	45.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.54%	
Target Compounds						
16) Acetone	2.45	43	12279	9.219	ug/l	99
20) Methylene Chloride	2.86	84	9647	4.473	ug/l	93
43) Isopropyl Acetate	6.45	43	14059	2.313	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050819\
Data File : VX009399.D
Acq On : 08 May 2019 15:33
Operator : JC/SP
Sample : PB119521TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB119521TB

Quant Time: May 09 06:56:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 03:31:13 2019
Response via : Initial Calibration

