

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050819\
 Data File : VX009401.D
 Acq On : 08 May 2019 16:20
 Operator : JC/SP
 Sample : K2641-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-02-050719-B

Quant Time: May 09 07:01:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 03:31:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	182406	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	291201	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	258258	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	109632	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	119331	47.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.84%	
35) Dibromofluoromethane	5.50	113	89705	49.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.00%	
50) Toluene-d8	8.71	98	370221	50.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.18%	
62) 4-Bromofluorobenzene	11.13	95	121626	45.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.58%	
Target Compounds						
16) Acetone	2.45	43	13493	10.092	ug/l	94
18) Methyl Acetate	2.78	43	5434	1.709	ug/l	98
20) Methylene Chloride	2.86	84	4521	2.088	ug/l	87
43) Isopropyl Acetate	6.45	43	11784	1.954	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050819\
Data File : VX009401.D
Acq On : 08 May 2019 16:20
Operator : JC/SP
Sample : K2641-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-02-050719-B

Quant Time: May 09 07:01:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 03:31:13 2019
Response via : Initial Calibration

