

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050819\  
 Data File : VX009404.D  
 Acq On : 08 May 2019 17:31  
 Operator : JC/SP  
 Sample : K2641-10  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 CL-02-050719-E

Quant Time: May 09 07:06:25 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X042919W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed May 08 03:31:13 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units  | Dev(Min) |
|-----------------------------|-------|------|----------|--------|--------|----------|
| 1) Pentafluorobenzene       | 5.67  | 168  | 176241   | 50.00  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.86  | 114  | 284176   | 50.00  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.12 | 117  | 246015   | 50.00  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.08 | 152  | 97586    | 50.00  | ug/l   | 0.00     |
| System Monitoring Compounds |       |      |          |        |        |          |
| 33) 1,2-Dichloroethane-d4   | 6.06  | 65   | 115387   | 47.96  | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 95.92% |          |
| 35) Dibromofluoromethane    | 5.50  | 113  | 85988    | 48.13  | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 96.26% |          |
| 50) Toluene-d8              | 8.71  | 98   | 357037   | 49.50  | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 99.00% |          |
| 62) 4-Bromofluorobenzene    | 11.13 | 95   | 113024   | 43.13  | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 86.26% |          |
| Target Compounds            |       |      |          |        |        |          |
| 16) Acetone                 | 2.45  | 43   | 28817    | 22.306 | ug/l   | 97       |
| 18) Methyl Acetate          | 2.78  | 43   | 5461     | 1.777  | ug/l   | 98       |
| 20) Methylene Chloride      | 2.85  | 84   | 57509    | 27.494 | ug/l   | 98       |
| 43) Isopropyl Acetate       | 6.45  | 43   | 12576    | 2.137  | ug/l   | 99       |
| 95) Naphthalene             | 13.83 | 128  | 105053   | 14.765 | ug/l   | 100      |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050819\  
Data File : VX009404.D  
Acq On : 08 May 2019 17:31  
Operator : JC/SP  
Sample : K2641-10  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 16 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
CL-02-050719-E

Quant Time: May 09 07:06:25 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X042919W.M  
Quant Title : SW846 8260  
QLast Update : Wed May 08 03:31:13 2019  
Response via : Initial Calibration

