

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050819\
 Data File : VX009405.D
 Acq On : 08 May 2019 17:54
 Operator : JC/SP
 Sample : K2643-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OR-03-050719-A

Quant Time: May 09 07:08:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 03:31:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	176505	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	285332	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	247426	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	97138	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	117466	48.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.50%	
35) Dibromofluoromethane	5.50	113	86979	48.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.98%	
50) Toluene-d8	8.71	98	359121	49.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.18%	
62) 4-Bromofluorobenzene	11.13	95	115137	43.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.52%	
Target Compounds						
16) Acetone	2.45	43	13863	10.715	ug/l	92
18) Methyl Acetate	2.78	43	4820	1.566	ug/l	92
20) Methylene Chloride	2.85	84	8264	3.945	ug/l	93
43) Isopropyl Acetate	6.45	43	12364	2.092	ug/l	99
95) Naphthalene	13.83	128	14047	1.983	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050819\
Data File : VX009405.D
Acq On : 08 May 2019 17:54
Operator : JC/SP
Sample : K2643-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OR-03-050719-A

Quant Time: May 09 07:08:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 03:31:13 2019
Response via : Initial Calibration

