

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050820\
 Data File : VX016124.D
 Acq On : 08 May 2020 16:26
 Operator : JC/SP
 Sample : L2552-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z4-TB2-200507

Quant Time: May 09 04:52:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	268857	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	447077	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	435921	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	229401	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	166631	47.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.20%	
35) Dibromofluoromethane	5.47	113	131393	46.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.78%	
50) Toluene-d8	8.70	98	548317	50.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.48%	
62) 4-Bromofluorobenzene	11.12	95	224298	54.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.06%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050820\
Data File : VX016124.D
Acq On : 08 May 2020 16:26
Operator : JC/SP
Sample : L2552-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
Z4-TB2-200507

Quant Time: May 09 04:52:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

