

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050820\
 Data File : VX016132.D
 Acq On : 08 May 2020 19:28
 Operator : JC/SP
 Sample : L2554-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DRUM-1-1

Quant Time: May 09 05:30:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	274230	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	451566	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	445975	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	237731	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	172252	47.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.46%	
35) Dibromofluoromethane	5.47	113	137515	48.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.14%	
50) Toluene-d8	8.70	98	559940	50.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.60%	
62) 4-Bromofluorobenzene	11.12	95	225050	53.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.36%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.00	59	4071	4.461	ug/l #	84
16) Acetone	2.42	43	64356	39.970	ug/l	98
18) Methyl Acetate	2.75	43	5636	1.424	ug/l	99
20) Methylene Chloride	2.84	84	15769	4.751	ug/l	98
25) 2-Butanone	4.64	43	8763	3.377	ug/l	96
29) Tetrahydrofuran	5.11	42	2361	1.373	ug/l	93
43) Isopropyl Acetate	6.42	43	98997	12.043	ug/l	99
95) Naphthalene	13.82	128	27418	1.482	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050820\
Data File : VX016132.D
Acq On : 08 May 2020 19:28
Operator : JC/SP
Sample : L2554-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DRUM-1-1

Quant Time: May 09 05:30:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

