

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050820\
 Data File : VX016134.D
 Acq On : 08 May 2020 20:13
 Operator : JC/SP
 Sample : L2557-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-13A

Quant Time: May 09 07:25:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	280582	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	470273	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	459440	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	238447	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	175471	47.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.06%	
35) Dibromofluoromethane	5.47	113	139961	46.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.96%	
50) Toluene-d8	8.70	98	584971	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
62) 4-Bromofluorobenzene	11.12	95	236376	54.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.28%	
Target Compounds						
16) Acetone	2.42	43	27853	16.907	ug/l	98
18) Methyl Acetate	2.75	43	4428m	1.093	ug/l	
20) Methylene Chloride	2.84	84	11562	3.404	ug/l	94
43) Isopropyl Acetate	6.42	43	84520	9.873	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050820\
Data File : VX016134.D
Acq On : 08 May 2020 20:13
Operator : JC/SP
Sample : L2557-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-13A

Quant Time: May 09 07:25:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

