

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050823\
 Data File : VX035553.D
 Acq On : 08 May 2023 12:26
 Operator : JC/MD
 Sample : VX0508WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBL01

Quant Time: May 09 03:40:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	266155	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	440786	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	403085	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	190678	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	156507	48.741	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.480%
35) Dibromofluoromethane	5.385	113	134671	49.778	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.560%
50) Toluene-d8	8.646	98	512110	50.179	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.360%
62) 4-Bromofluorobenzene	11.079	95	197707	51.497	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050823\
Data File : VX035553.D
Acq On : 08 May 2023 12:26
Operator : JC/MD
Sample : VX0508WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBL01

Quant Time: May 09 03:40:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

