

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050823\
 Data File : VX035557.D
 Acq On : 08 May 2023 14:04
 Operator : JC/MD
 Sample : 02643-09DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-PRO-06-99-05042023DL

Quant Time: May 09 03:43:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	246642	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	402085	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	374806	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	178097	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	146198	49.132	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.260%
35) Dibromofluoromethane	5.385	113	126280	51.169	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.340%
50) Toluene-d8	8.647	98	470535	50.543	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.080%
62) 4-Bromofluorobenzene	11.079	95	184165	52.587	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.180%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	3451	1.441	ug/l	94
20) Methylene Chloride	2.788	84	1990	0.672	ug/l #	91
44) Trichloroethene	7.129	130	156486	51.387	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050823\
Data File : VX035557.D
Acq On : 08 May 2023 14:04
Operator : JC/MD
Sample : 02643-09DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-PRO-06-99-05042023DL

Quant Time: May 09 03:43:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

