

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050823\
 Data File : VX035565.D
 Acq On : 08 May 2023 17:11
 Operator : JC/MD
 Sample : 02660-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WATER-SAMPLE-API-3-5

Quant Time: May 09 03:46:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	235862	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	388843	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	353179	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	168444	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	143267	50.348	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.700%	
35) Dibromofluoromethane	5.385	113	119871	50.227	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.460%	
50) Toluene-d8	8.647	98	456012	50.651	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.300%	
62) 4-Bromofluorobenzene	11.079	95	174372	51.486	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.980%	
Target Compounds						
16) Acetone	2.386	43	5990	4.847	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050823\
Data File : VX035565.D
Acq On : 08 May 2023 17:11
Operator : JC/MD
Sample : 02660-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WATER-SAMPLE-API-3-5

Quant Time: May 09 03:46:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

