

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050823\
 Data File : VX035572.D
 Acq On : 08 May 2023 19:54
 Operator : JC/MD
 Sample : 02648-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ORA-1635

Quant Time: May 09 03:51:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	269086	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	438355	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	362653	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	194821	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	159767	49.214	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.420%
35) Dibromofluoromethane	5.391	113	138799	51.589	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.180%
50) Toluene-d8	8.653	98	517332	50.972	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.940%
62) 4-Bromofluorobenzene	11.085	95	198726	52.049	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.100%
Target Compounds						
					Qvalue	
17) Carbon Disulfide	2.513	76	3259	0.520	ug/l	96
30) Chloroform	5.098	83	4610	0.891	ug/l	98
52) Toluene	8.720	92	11713	1.533	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	21378	1.772	ug/l	97
95) Naphthalene	13.774	128	27958	2.147	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050823\
Data File : VX035572.D
Acq On : 08 May 2023 19:54
Operator : JC/MD
Sample : 02648-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ORA-1635

Quant Time: May 09 03:51:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

