


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050823\
Data File : VX035550.D
Acq On : 08 May 2023 09:24
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Title : SW846 8260
Last Update : Tue Apr 25 16:31:40 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	17.0	16279	PASS
75	95	30	60	49.6	47355	PASS
95	95	100	100	100.0	95549	PASS
96	95	5	9	7.0	6717	PASS
173	174	0.00	2	0.6	421	PASS
174	95	50	100	75.5	72171	PASS
175	174	5	9	7.5	5385	PASS
176	174	95	101	97.0	70003	PASS
177	176	5	9	6.4	4488	PASS