

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050824\
 Data File : VX041415.D
 Acq On : 08 May 2024 11:46
 Operator : JC/MD
 Sample : VX0508WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBL01

Quant Time: May 09 05:30:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	174445	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	254234	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240960	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	114969	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	93816	49.401	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.800%
35) Dibromofluoromethane	5.385	113	77104	44.753	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.500%
50) Toluene-d8	8.647	98	281300	51.586	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.180%
62) 4-Bromofluorobenzene	11.079	95	93967	44.515	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.040%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050824\
Data File : VX041415.D
Acq On : 08 May 2024 11:46
Operator : JC/MD
Sample : VX0508WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBL01

Quant Time: May 09 05:30:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

