

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050824\
 Data File : VX041418.D
 Acq On : 08 May 2024 14:06
 Operator : JC/MD
 Sample : P2389-02 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1401

Quant Time: May 09 05:32:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	178875	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	306352	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	253716	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	152669	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	99933	51.319	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 102.640%		
35) Dibromofluoromethane	5.385	113	80677	38.861	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 77.720%		
50) Toluene-d8	8.647	98	339877	51.725	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 103.440%		
62) 4-Bromofluorobenzene	11.079	95	131561	51.722	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 103.440%		
Target Compounds						
						Qvalue
16) Acetone	2.386	43	6654	3.909	ug/l	100
18) Methyl Acetate	2.709	43	4765	2.392	ug/l	98
20) Methylene Chloride	2.788	84	1148	0.679	ug/l	97
40) Benzene	6.044	78	2776	0.363	ug/l	100
52) Toluene	8.726	92	2671	0.571	ug/l	100
70) Styrene	10.659	104	5047	0.944	ug/l	90
95) Naphthalene	13.780	128	7550	0.722	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050824\
Data File : VX041418.D
Acq On : 08 May 2024 14:06
Operator : JC/MD
Sample : P2389-02 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1401

Quant Time: May 09 05:32:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

