

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050824\
 Data File : VX041422.D
 Acq On : 08 May 2024 16:42
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 09 05:34:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	78	0.00
2 T	Dichlorodifluoromethane	0.411	0.401	2.4	66	0.00
3 P	Chloromethane	0.425	0.394	7.3	73	0.00
4 C	Vinyl Chloride	0.405	0.420	-3.7#	80	0.00
5 T	Bromomethane	0.254	0.277	-9.1	83	0.00
6 T	Chloroethane	0.188	0.222	-18.1	92	0.00
7 T	Trichlorofluoromethane	0.699	0.673	3.7	78	0.00
8 T	Diethyl Ether	0.246	0.283	-15.0	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.393	0.445	-13.2	84	0.00
10 T	Methyl Iodide	0.423	0.527	-24.6	88	0.00
11 T	Tert butyl alcohol	0.109	0.128	-17.4	91	-0.01
12 CM	1,1-Dichloroethene	0.387	0.440	-13.7#	84	0.00
13 T	Acrolein	0.086	0.104	-20.9	97	0.00
14 T	Allyl chloride	0.651	0.769	-18.1	86	0.00
15 T	Acrylonitrile	0.249	0.302	-21.3	96	0.00
16 T	Acetone	0.284	0.269	5.3	91	0.00
17 T	Carbon Disulfide	0.911	0.896	1.6	72	0.00
18 T	Methyl Acetate	0.557	0.703	-26.2#	103	0.00
19 T	Methyl tert-butyl Ether	1.412	1.641	-16.2	92	0.00
20 T	Methylene Chloride	0.473	0.510	-7.8	89	0.00
21 T	trans-1,2-Dichloroethene	0.433	0.459	-6.0	84	0.00
22 T	Diisopropyl ether	1.380	1.591	-15.3	91	0.00
23 T	Vinyl Acetate	1.234	1.447	-17.3	89	0.00
24 P	1,1-Dichloroethane	0.779	0.844	-8.3	86	0.00
25 T	2-Butanone	0.352	0.437	-24.1	96	0.00
26 T	2,2-Dichloropropane	0.652	0.632	3.1	75	0.00
27 T	cis-1,2-Dichloroethene	0.513	0.580	-13.1	87	0.00
28 T	Bromochloromethane	0.278	0.296	-6.5	90	0.00
29 T	Tetrahydrofuran	0.230	0.281	-22.2	98	0.00
30 C	Chloroform	0.811	0.939	-15.8#	91	0.00
31 T	Cyclohexane	0.657	0.618	5.9	74	0.00
32 T	1,1,1-Trichloroethane	0.738	0.815	-10.4	87	0.00
33 S	1,2-Dichloroethane-d4	0.544	0.625	-14.9	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.339	0.347	-2.4	88	0.00
36 T	1,1-Dichloropropene	0.405	0.386	4.7	81	0.00
37 T	Ethyl Acetate	0.506	0.571	-12.8	95	0.00
38 T	Carbon Tetrachloride	0.464	0.462	0.4	82	0.00
39 T	Methylcyclohexane	0.476	0.470	1.3	91	0.00
40 TM	Benzene	1.249	1.287	-3.0	90	0.00
41 T	Methacrylonitrile	0.260	0.302	-16.2	93	0.00
42 TM	1,2-Dichloroethane	0.458	0.505	-10.3	94	0.00
43 T	Isopropyl Acetate	0.760	0.875	-15.1	98	0.00
44 TM	Trichloroethene	0.330	0.353	-7.0	102	0.00
45 C	1,2-Dichloropropane	0.279	0.321	-15.1#	105	0.00
46 T	Dibromomethane	0.222	0.253	-14.0	99	0.00
47 T	Bromodichloromethane	0.396	0.484	-22.2	99	0.00
48 T	Methyl methacrylate	0.348	0.431	-23.9	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050824\
 Data File : VX041422.D
 Acq On : 08 May 2024 16:42
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 09 05:34:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.009	-12.5	102	0.00
50 S	Toluene-d8	1.072	1.166	-8.8	96	0.00
51 T	4-Methyl-2-Pentanone	0.456	0.589	-29.2#	107	0.00
52 CM	Toluene	0.764	0.825	-8.0#	95	0.00
53 T	t-1,3-Dichloropropene	0.392	0.421	-7.4	87	0.00
54 T	cis-1,3-Dichloropropene	0.435	0.506	-16.3	94	0.00
55 T	1,1,2-Trichloroethane	0.294	0.309	-5.1	91	0.00
56 T	Ethyl methacrylate	0.455	0.496	-9.0	88	0.00
57 T	1,3-Dichloropropane	0.491	0.511	-4.1	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.221	0.275	-24.4	101	0.00
59 T	2-Hexanone	0.365	0.423	-15.9	98	0.00
60 T	Dibromochloromethane	0.346	0.376	-8.7	88	0.00
61 T	1,2-Dibromoethane	0.324	0.332	-2.5	91	0.00
62 S	4-Bromofluorobenzene	0.415	0.376	9.4	76	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	77	0.00
64 T	Tetrachloroethene	0.362	0.386	-6.6	83	0.00
65 PM	Chlorobenzene	1.015	1.061	-4.5	84	0.00
66 T	1,1,1,2-Tetrachloroethane	0.359	0.396	-10.3	81	0.00
67 C	Ethyl Benzene	1.675	1.707	-1.9#	78	0.00
68 T	m/p-Xylenes	0.671	0.685	-2.1	78	0.00
69 T	o-Xylene	0.642	0.689	-7.3	76	0.00
70 T	Styrene	1.054	1.183	-12.2	77	0.00
71 P	Bromoform	0.321	0.384	-19.6	82	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
73 T	Isopropylbenzene	3.052	3.055	-0.1	80	0.00
74 T	N-amyl acetate	1.333	1.576	-18.2	84	0.00
75 P	1,1,2,2-Tetrachloroethane	1.000	1.149	-14.9	89	0.00
76 T	1,2,3-Trichloropropane	0.815	0.932	-14.4	93	0.00
77 T	Bromobenzene	0.861	0.921	-7.0	82	0.00
78 T	n-propylbenzene	3.206	3.375	-5.3	84	0.00
79 T	2-Chlorotoluene	2.066	2.127	-3.0	85	0.00
80 T	1,3,5-Trimethylbenzene	2.488	2.593	-4.2	78	0.00
81 T	trans-1,4-Dichloro-2-butene	0.299	0.336	-12.4	81	0.00
82 T	4-Chlorotoluene	2.345	2.445	-4.3	79	0.00
83 T	tert-Butylbenzene	2.664	2.798	-5.0	80	0.00
84 T	1,2,4-Trimethylbenzene	2.503	2.668	-6.6	83	0.00
85 T	sec-Butylbenzene	3.053	3.256	-6.6	79	0.00
86 T	p-Isopropyltoluene	2.742	2.914	-6.3	80	0.00
87 T	1,3-Dichlorobenzene	1.566	1.643	-4.9	80	0.00
88 T	1,4-Dichlorobenzene	1.613	1.644	-1.9	80	0.00
89 T	n-Butylbenzene	2.205	2.336	-5.9	79	0.00
90 T	Hexachloroethane	0.459	0.489	-6.5	77	0.00
91 T	1,2-Dichlorobenzene	1.654	1.730	-4.6	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.231	0.273	-18.2	86	0.00
93 T	1,2,4-Trichlorobenzene	1.156	1.223	-5.8	77	0.00
94 T	Hexachlorobutadiene	0.577	0.567	1.7	77	0.00
95 T	Naphthalene	3.426	3.935	-14.9	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050824\
Data File : VX041422.D
Acq On : 08 May 2024 16:42
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 09 05:34:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.181	1.300	-10.1	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6