

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046084.D
 Acq On : 08 May 2025 12:32
 Operator : JC/MD
 Sample : Q1985-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-20250507

Quant Time: May 09 01:56:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	64053	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	126115	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	114932	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49527	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61356	51.380	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.760%
35) Dibromofluoromethane	5.385	113	44864	49.401	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.800%
50) Toluene-d8	8.647	98	154594	49.183	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.360%
62) 4-Bromofluorobenzene	11.079	95	59587	49.421	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046084.D
Acq On : 08 May 2025 12:32
Operator : JC/MD
Sample : Q1985-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20250507

Quant Time: May 09 01:56:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

