

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046093.D
 Acq On : 08 May 2025 16:02
 Operator : JC/MD
 Sample : Q1972-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-B-7

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/09/2025
 Supervised By : Mahesh Dadoda 05/09/2025

Quant Time: May 09 01:58:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	62561	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	126690	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118927	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	50040	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62401	53.502	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.000%			
35) Dibromofluoromethane	5.379	113	45570	49.951	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.900%			
50) Toluene-d8	8.647	98	157544	49.894	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.780%			
62) 4-Bromofluorobenzene	11.079	95	60063	49.589	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 99.180%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	22575	48.274	ug/l	99
18) Methyl Acetate	2.709	43	2134	1.966	ug/l	98
20) Methylene Chloride	2.782	84	2116	2.361	ug/l	94
25) 2-Butanone	4.599	43	3822m	5.630	ug/l	
43) Isopropyl Acetate	6.342	43	11504	4.979	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046093.D
Acq On : 08 May 2025 16:02
Operator : JC/MD
Sample : Q1972-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-B-7

Quant Time: May 09 01:58:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025

