

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046094.D
 Acq On : 08 May 2025 16:26
 Operator : JC/MD
 Sample : Q1972-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SUB-WC-2

Quant Time: May 09 01:58:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	61992	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	121402	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	116858	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	47611	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61340	53.075	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 106.140%		
35) Dibromofluoromethane	5.379	113	44540	50.948	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.900%		
50) Toluene-d8	8.647	98	156852	51.838	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.680%		
62) 4-Bromofluorobenzene	11.079	95	59463	51.232	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 102.460%		
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	21880	47.217	ug/l	99
18) Methyl Acetate	2.703	43	2214	2.059	ug/l	93
20) Methylene Chloride	2.782	84	2244	2.527	ug/l	86
43) Isopropyl Acetate	6.342	43	12006	5.423	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046094.D
Acq On : 08 May 2025 16:26
Operator : JC/MD
Sample : Q1972-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SUB-WC-2

Quant Time: May 09 01:58:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

