

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046095.D
 Acq On : 08 May 2025 16:49
 Operator : JC/MD
 Sample : Q1972-24
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-B-1

Quant Time: May 09 01:58:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	63689	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	127040	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	120467	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51286	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64746	54.529	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 109.060%		
35) Dibromofluoromethane	5.385	113	46532	50.865	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.720%		
50) Toluene-d8	8.647	98	161352	50.959	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 101.920%		
62) 4-Bromofluorobenzene	11.079	95	62126	51.151	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 102.300%		
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	22638	47.551	ug/l	98
18) Methyl Acetate	2.703	43	2444	2.212	ug/l	98
20) Methylene Chloride	2.782	84	2093	2.294	ug/l	84
43) Isopropyl Acetate	6.348	43	12256	5.290	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046095.D
Acq On : 08 May 2025 16:49
Operator : JC/MD
Sample : Q1972-24
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-B-1

Quant Time: May 09 01:58:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

