

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050919\
 Data File : VX009424.D
 Acq On : 09 May 2019 13:51
 Operator : JC/SP
 Sample : K2748-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7

Quant Time: May 10 07:40:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	186577	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	299302	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	258221	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	106801	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	121859	47.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.68%	
35) Dibromofluoromethane	5.50	113	91644	48.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.42%	
50) Toluene-d8	8.71	98	372660	49.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.12%	
62) 4-Bromofluorobenzene	11.13	95	118428	42.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.82%	
Target Compounds						
16) Acetone	2.44	43	4735	3.462	ug/l	97
19) Methyl tert-butyl Ether	3.20	73	49948	7.347	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050919\
Data File : VX009424.D
Acq On : 09 May 2019 13:51
Operator : JC/SP
Sample : K2748-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7

Quant Time: May 10 07:40:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

