

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050919\
 Data File : VX009428.D
 Acq On : 09 May 2019 15:25
 Operator : JC/SP
 Sample : K2748-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-12

Quant Time: May 10 07:48:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	183719	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	296648	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	255740	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	104009	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	121794	48.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.12%	
35) Dibromofluoromethane	5.50	113	90518	48.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.08%	
50) Toluene-d8	8.71	98	366812	48.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.44%	
62) 4-Bromofluorobenzene	11.13	95	119082	43.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.06%	
Target Compounds						
16) Acetone	2.44	43	4836	3.591	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050919\
Data File : VX009428.D
Acq On : 09 May 2019 15:25
Operator : JC/SP
Sample : K2748-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-12

Quant Time: May 10 07:48:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

