

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050919\
 Data File : VX009437.D
 Acq On : 09 May 2019 18:54
 Operator : JC/SP
 Sample : K2786-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB202-GW-218-220

Quant Time: May 10 08:06:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	188120	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	302200	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	262265	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	116010	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	122979	47.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.78%	
35) Dibromofluoromethane	5.50	113	92064	48.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.92%	
50) Toluene-d8	8.71	98	374745	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
62) 4-Bromofluorobenzene	11.13	95	125138	44.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.80%	
Target Compounds						
					Qvalue	
7) Trichlorofluoromethane	1.93	101	4540	1.618	ug/l	94
12) 1,1-Dichloroethene	2.37	96	723	0.410	ug/l	85
16) Acetone	2.44	43	18387	13.334	ug/l	99
19) Methyl tert-butyl Ether	3.19	73	5693	0.831	ug/l	97
25) 2-Butanone	4.68	43	4420	2.063	ug/l #	87
44) Trichloroethene	7.21	130	1553	0.739	ug/l	96
64) Tetrachloroethene	9.34	164	4424	2.459	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050919\
Data File : VX009437.D
Acq On : 09 May 2019 18:54
Operator : JC/SP
Sample : K2786-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB202-GW-218-220

Quant Time: May 10 08:06:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

