

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050919\
 Data File : VX009438.D
 Acq On : 09 May 2019 19:18
 Operator : JC/SP
 Sample : K2786-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB202-GW-238-240

Quant Time: May 10 08:08:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	186940	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	299932	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	265775	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	116895	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	123262	48.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.60%	
35) Dibromofluoromethane	5.50	113	92237	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
50) Toluene-d8	8.71	98	376957	49.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.04%	
62) 4-Bromofluorobenzene	11.13	95	126585	45.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.52%	
Target Compounds						
7) Trichlorofluoromethane	1.93	101	6204	2.225	ug/l	98
12) 1,1-Dichloroethene	2.38	96	654	0.373	ug/l	89
16) Acetone	2.44	43	8314	6.067	ug/l	96
30) Chloroform	5.21	83	2243	0.612	ug/l	90
44) Trichloroethene	7.22	130	1639	0.786	ug/l	99
64) Tetrachloroethene	9.34	164	3336	1.830	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050919\
Data File : VX009438.D
Acq On : 09 May 2019 19:18
Operator : JC/SP
Sample : K2786-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB202-GW-238-240

Quant Time: May 10 08:08:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

