

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050919\
 Data File : VX009452.D
 Acq On : 10 May 2019 01:30
 Operator : JC/SP
 Sample : K2782-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW272-190507

Quant Time: May 10 08:37:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	196367	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	312093	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	273243	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	117895	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	127331	47.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.00%	
35) Dibromofluoromethane	5.50	113	94174	48.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.00%	
50) Toluene-d8	8.71	98	386919	48.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.70%	
62) 4-Bromofluorobenzene	11.13	95	128497	44.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.30%	
Target Compounds						
19) Methyl tert-butyl Ether	3.19	73	4551	0.636	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	6264	2.626	ug/l	89
44) Trichloroethene	7.21	130	25208	11.617	ug/l	97
64) Tetrachloroethene	9.34	164	7463	3.982	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050919\
Data File : VX009452.D
Acq On : 10 May 2019 01:30
Operator : JC/SP
Sample : K2782-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW272-190507

Quant Time: May 10 08:37:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

