

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050919\
 Data File : VX009453.D
 Acq On : 10 May 2019 01:53
 Operator : JC/SP
 Sample : K2782-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW271-190507

Quant Time: May 10 08:38:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	188404	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	303399	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	264607	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	113581	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	124810	48.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.06%	
35) Dibromofluoromethane	5.50	113	92527	48.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.02%	
50) Toluene-d8	8.71	98	380746	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
62) 4-Bromofluorobenzene	11.13	95	124825	44.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.22%	
Target Compounds						
19) Methyl tert-butyl Ether	3.19	73	4627	0.674	ug/l	98
27) cis-1,2-Dichloroethene	4.59	96	2633	1.151	ug/l	92
44) Trichloroethene	7.21	130	8236	3.904	ug/l	97
64) Tetrachloroethene	9.34	164	2831	1.560	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050919\
Data File : VX009453.D
Acq On : 10 May 2019 01:53
Operator : JC/SP
Sample : K2782-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW271-190507

Quant Time: May 10 08:38:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

