

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050919\
 Data File : VX009463.D
 Acq On : 10 May 2019 05:46
 Operator : JC/SP
 Sample : K2782-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW210-190507

Quant Time: May 10 09:02:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	184852	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	298975	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	258717	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	111388	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	122311	48.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.94%	
35) Dibromofluoromethane	5.50	113	90953	48.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.78%	
50) Toluene-d8	8.71	98	368493	48.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.12%	
62) 4-Bromofluorobenzene	11.13	95	120370	43.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.32%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	4013	1.787	ug/l	89
30) Chloroform	5.21	83	1245	0.343	ug/l #	84
44) Trichloroethene	7.21	130	21022	10.113	ug/l	99
64) Tetrachloroethene	9.34	164	4068	2.292	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050919\
Data File : VX009463.D
Acq On : 10 May 2019 05:46
Operator : JC/SP
Sample : K2782-18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW210-190507

Quant Time: May 10 09:02:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

