

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050922\
 Data File : VX028568.D
 Acq On : 09 May 2022 13:40
 Operator : JC/MD
 Sample : N2650-01
 Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ETGI-349

Quant Time: May 09 15:51:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	465729	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	756404	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	691435	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	341102	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	251835	41.471	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	82.940%		
35) Dibromofluoromethane	5.391	113	231148	43.575	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	87.140%		
50) Toluene-d8	8.653	98	863122	42.364	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	84.720%		
62) 4-Bromofluorobenzene	11.085	95	309099	39.678	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	79.360%		
Target Compounds						
25) 2-Butanone	4.611	43	29788	8.283	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050922\
Data File : VX028568.D
Acq On : 09 May 2022 13:40
Operator : JC/MD
Sample : N2650-01
Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ETGI-349

Quant Time: May 09 15:51:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

