



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050922\
 Data File : VX028561.D
 Acq On : 09 May 2022 09:17
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Title : SW846 8260
 Last Update : Tue Apr 19 13:38:04 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.3	15166	PASS
75	95	30	60	50.8	42141	PASS
95	95	100	100	100.0	82981	PASS
96	95	5	9	6.6	5490	PASS
173	174	0.00	2	1.1	747	PASS
174	95	50	100	81.0	67248	PASS
175	174	5	9	8.0	5382	PASS
176	174	95	101	96.7	65000	PASS
177	176	5	9	6.6	4320	PASS