

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050924\
 Data File : VX041432.D
 Acq On : 09 May 2024 14:09
 Operator : JC/MD
 Sample : VX0509MBL01
 Misc : 5.00g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: May 09 23:39:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML050924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu May 09 23:24:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	299975	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	270179	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	138876	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	87370	51.354	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 102.700%	
7) Chloroethane-d5	1.648	69	62630	50.175	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 100.360%	
11) 1,1-Dichloroethene-d2	2.294	65	39710	46.375	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 92.760%	
21) 2-Butanone-d5	4.459	46	152812	120.478	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 120.480%	
24) Chloroform-d	5.050	84	195406	51.940	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 103.880%	
26) 1,2-Dichloroethane-d4	5.952	65	136203	50.326	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 100.660%	
32) Benzene-d6	5.964	84	403133	54.618	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 109.240%	
36) 1,2-Dichloropropane-d6	7.306	67	116803	53.258	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 106.520%	
41) Toluene-d8	8.647	98	373701	49.020	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 98.040%	
43) trans-1,3-Dichloroprop...	8.952	79	53589	48.773	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 97.540%	
47) 2-Hexanone-d5	9.385	63	112610	110.191	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 110.190%	
56) 1,1,2,2-Tetrachloroeth...	11.189	84	174792	50.644	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 101.280%	
66) 1,2-Dichlorobenzene-d4	12.317	152	158395	55.819	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 111.640%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050924\
Data File : VX041432.D
Acq On : 09 May 2024 14:09
Operator : JC/MD
Sample : VX0509MBL01
Misc : 5.00g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: May 09 23:39:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML050924WMA.M
Quant Title : VOC Analysis
QLast Update : Thu May 09 23:24:52 2024
Response via : Initial Calibration

