

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051019\
 Data File : VX009472.D
 Acq On : 10 May 2019 10:44
 Operator : JC/SP
 Sample : VX0510WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0510WBL01

Quant Time: May 10 15:15:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X050119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 02 06:46:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.01	128	29050	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.85	114	165699	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	144467	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.06	65	70566	29.28	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	97.60%
60) 4-Bromofluorobenzene	11.13	95	68065	27.73	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	92.43%
63) Toluene-d8	8.71	98	205604	30.02	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	100.07%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051019\
Data File : VX009472.D
Acq On : 10 May 2019 10:44
Operator : JC/SP
Sample : VX0510WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0510WBL01

Quant Time: May 10 15:15:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X050119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 02 06:46:57 2019
Response via : Initial Calibration

