

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051019\
 Data File : VX009495.D
 Acq On : 10 May 2019 20:50
 Operator : JC/SP
 Sample : K2783-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW720-190508

Quant Time: May 11 04:04:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	180099	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	288431	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	249265	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	104921	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	116178	47.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.50%	
35) Dibromofluoromethane	5.50	113	87946	48.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.00%	
50) Toluene-d8	8.71	98	356780	48.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.48%	
62) 4-Bromofluorobenzene	11.13	95	116508	43.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.60%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	1243	0.568	ug/l	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051019\
Data File : VX009495.D
Acq On : 10 May 2019 20:50
Operator : JC/SP
Sample : K2783-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW720-190508

Quant Time: May 11 04:04:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

