

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051019\
 Data File : VX009496.D
 Acq On : 10 May 2019 21:13
 Operator : JC/SP
 Sample : K2783-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW721-190508

Quant Time: May 11 04:06:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	179387	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	286751	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	249734	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	105178	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	118135	48.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.48%	
35) Dibromofluoromethane	5.50	113	87012	48.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.54%	
50) Toluene-d8	8.71	98	360841	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
62) 4-Bromofluorobenzene	11.13	95	117504	44.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.86%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051019\
Data File : VX009496.D
Acq On : 10 May 2019 21:13
Operator : JC/SP
Sample : K2783-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW721-190508

Quant Time: May 11 04:06:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

