

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051019\
 Data File : VX009498.D
 Acq On : 10 May 2019 22:00
 Operator : JC/SP
 Sample : K2783-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW215-190508

Quant Time: May 11 04:09:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	182607	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	291597	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	253118	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	105986	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	120616	48.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.78%	
35) Dibromofluoromethane	5.50	113	88926	48.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.02%	
50) Toluene-d8	8.71	98	363091	49.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.12%	
62) 4-Bromofluorobenzene	11.13	95	117714	43.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.54%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	15374	6.932	ug/l	89
44) Trichloroethene	7.21	130	9945	4.905	ug/l	98
64) Tetrachloroethene	9.34	164	3249	1.871	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051019\
Data File : VX009498.D
Acq On : 10 May 2019 22:00
Operator : JC/SP
Sample : K2783-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW215-190508

Quant Time: May 11 04:09:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

