

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051019\
 Data File : VX009500.D
 Acq On : 10 May 2019 22:47
 Operator : JC/SP
 Sample : K2781-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-WK-2

Quant Time: May 11 04:14:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	178060	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	282427	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	238807	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	93039	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	115469	47.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.00%	
35) Dibromofluoromethane	5.50	113	86382	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
50) Toluene-d8	8.71	98	349304	48.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.46%	
62) 4-Bromofluorobenzene	11.13	95	107935	41.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.88%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	26531	12.268	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051019\
Data File : VX009500.D
Acq On : 10 May 2019 22:47
Operator : JC/SP
Sample : K2781-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-WK-2

Quant Time: May 11 04:14:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

