

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051019\
 Data File : VX009503.D
 Acq On : 10 May 2019 23:57
 Operator : JC/SP
 Sample : K2810-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW764-190509

Quant Time: May 11 04:26:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	179782	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	289665	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	253774	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	108252	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	118125	48.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.26%	
35) Dibromofluoromethane	5.50	113	88187	48.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.86%	
50) Toluene-d8	8.71	98	363504	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
62) 4-Bromofluorobenzene	11.13	95	119205	44.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.24%	
Target Compounds						
19) Methyl tert-butyl Ether	3.20	73	3344	0.510	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	7842	3.591	ug/l	88
44) Trichloroethene	7.21	130	30139	14.965	ug/l	96
64) Tetrachloroethene	9.34	164	5664	3.254	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051019\
Data File : VX009503.D
Acq On : 10 May 2019 23:57
Operator : JC/SP
Sample : K2810-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW764-190509

Quant Time: May 11 04:26:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

