

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051019\
 Data File : VX009504.D
 Acq On : 11 May 2019 00:20
 Operator : JC/SP
 Sample : K2810-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS004MW153-190509

Quant Time: May 11 04:27:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	179792	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	287535	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	249193	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	100748	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	117866	48.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.04%	
35) Dibromofluoromethane	5.50	113	87803	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
50) Toluene-d8	8.71	98	358851	49.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.34%	
62) 4-Bromofluorobenzene	11.13	95	115631	43.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.22%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	3151	1.443	ug/l	87
30) Chloroform	5.22	83	4612	1.307	ug/l	91
44) Trichloroethene	7.21	130	31697	15.855	ug/l	99
64) Tetrachloroethene	9.34	164	1074	0.628	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051019\
Data File : VX009504.D
Acq On : 11 May 2019 00:20
Operator : JC/SP
Sample : K2810-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS004MW153-190509

Quant Time: May 11 04:27:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

