

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051019\
 Data File : VX009508.D
 Acq On : 11 May 2019 01:54
 Operator : JC/SP
 Sample : K2805-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: May 11 04:34:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	183335	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	292321	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	255480	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	109796	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	120847	48.29	ug/l	0.00
Spiked Amount			Recovery	=	96.58%	
35) Dibromofluoromethane	5.50	113	89828	48.88	ug/l	0.00
Spiked Amount			Recovery	=	97.76%	
50) Toluene-d8	8.71	98	364114	49.08	ug/l	0.00
Spiked Amount			Recovery	=	98.16%	
62) 4-Bromofluorobenzene	11.13	95	120929	44.86	ug/l	0.00
Spiked Amount			Recovery	=	89.72%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051019\
Data File : VX009508.D
Acq On : 11 May 2019 01:54
Operator : JC/SP
Sample : K2805-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: May 11 04:34:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

