

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX051019\
 Data File : VX009511.D
 Acq On : 11 May 2019 03:04
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 11 04:41:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	78	0.00
2 T	Dichlorodifluoromethane	0.461	0.443	3.9	77	0.00
3 P	Chloromethane	0.629	0.651	-3.5	84	0.00
4 C	Vinyl Chloride	0.613	0.658	-7.3#	85	0.00
5 T	Bromomethane	0.368	0.375	-1.9	85	0.00
6 T	Chloroethane	0.394	0.414	-5.1	85	0.00
7 T	Trichlorofluoromethane	0.746	0.804	-7.8	86	0.00
8 T	Diethyl Ether	0.353	0.377	-6.8	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.468	0.484	-3.4	83	0.00
10 T	Methyl Iodide	0.483	0.605	-25.3#	93	0.00
11 T	Tert butyl alcohol	0.158	0.175	-10.8	89	0.00
12 CM	1,1-Dichloroethene	0.469	0.507	-8.1#	85	0.00
13 T	Acrolein	0.143	0.147	-2.8	83	0.00
14 T	Allyl chloride	1.143	1.173	-2.6	83	0.00
15 T	Acrylonitrile	0.386	0.418	-8.3	89	0.00
16 T	Acetone	0.367	0.364	0.8	81	0.00
17 T	Carbon Disulfide	1.317	1.436	-9.0	85	0.00
18 T	Methyl Acetate	0.872	0.904	-3.7	87	0.00
19 T	Methyl tert-butyl Ether	1.822	1.965	-7.8	86	0.00
20 T	Methylene Chloride	0.593	0.622	-4.9	87	0.00
21 T	trans-1,2-Dichloroethene	0.509	0.542	-6.5	85	0.00
22 T	Diisopropyl ether	2.199	2.377	-8.1	87	0.00
23 T	Vinyl Acetate	1.883	2.119	-12.5	87	0.00
24 P	1,1-Dichloroethane	1.085	1.147	-5.7	86	0.00
25 T	2-Butanone	0.570	0.618	-8.4	87	0.00
26 T	2,2-Dichloropropane	0.807	0.608	24.7#	60	0.00
27 T	cis-1,2-Dichloroethene	0.607	0.655	-7.9	88	0.00
28 T	Bromochloromethane	0.487	0.434	10.9	69	0.00
29 T	Tetrahydrofuran	0.366	0.415	-13.4	90	0.00
30 C	Chloroform	0.981	1.067	-8.8#	88	0.00
31 T	Cyclohexane	1.009	1.104	-9.4	86	0.00
32 T	1,1,1-Trichloroethane	0.817	0.879	-7.6	87	0.00
33 S	1,2-Dichloroethane-d4	0.683	0.682	0.1	78	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	79	0.00
35 S	Dibromofluoromethane	0.314	0.311	1.0	78	0.00
36 T	1,1-Dichloropropene	0.465	0.488	-4.9	86	0.00
37 T	Ethyl Acetate	0.647	0.718	-11.0	90	0.00
38 T	Carbon Tetrachloride	0.422	0.450	-6.6	86	0.00
39 T	Methylcyclohexane	0.571	0.591	-3.5	83	0.00
40 TM	Benzene	1.438	1.540	-7.1	87	0.00
41 T	Methacrylonitrile	0.385	0.413	-7.3	89	0.00
42 TM	1,2-Dichloroethane	0.516	0.555	-7.6	89	0.00
43 T	Isopropyl Acetate	1.036	1.128	-8.9	88	0.00
44 TM	Trichloroethene	0.348	0.361	-3.7	86	0.00
45 C	1,2-Dichloropropane	0.407	0.433	-6.4#	86	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX051019\
 Data File : VX009511.D
 Acq On : 11 May 2019 03:04
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 11 04:41:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.236	0.250	-5.9	88	0.00
47 T	Bromodichloromethane	0.472	0.503	-6.6	86	0.00
48 T	Methyl methacrylate	0.514	0.575	-11.9	90	0.00
49 T	1,4-Dioxane	0.010	0.011	-10.0	90	0.00
50 S	Toluene-d8	1.269	1.236	2.6	76	0.00
51 T	4-Methyl-2-Pentanone	0.677	0.757	-11.8	90	0.00
52 CM	Toluene	0.874	0.933	-6.8#	87	0.00
53 T	t-1,3-Dichloropropene	0.533	0.551	-3.4	81	0.00
54 T	cis-1,3-Dichloropropene	0.589	0.620	-5.3	82	0.00
55 T	1,1,2-Trichloroethane	0.357	0.381	-6.7	87	0.00
56 T	Ethyl methacrylate	0.637	0.696	-9.3	87	0.00
57 T	1,3-Dichloropropane	0.628	0.669	-6.5	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.338	0.329	2.7	75	0.00
59 T	2-Hexanone	0.532	0.590	-10.9	89	0.00
60 T	Dibromochloromethane	0.348	0.382	-9.8	87	0.00
61 T	1,2-Dibromoethane	0.360	0.392	-8.9	88	0.00
62 S	4-Bromofluorobenzene	0.461	0.456	1.1	77	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	80	0.00
64 T	Tetrachloroethene	0.343	0.378	-10.2	91	0.00
65 PM	Chlorobenzene	1.000	1.066	-6.6	87	0.00
66 T	1,1,1,2-Tetrachloroethane	0.363	0.392	-8.0	88	0.00
67 C	Ethyl Benzene	1.841	1.979	-7.5#	87	0.00
68 T	m/p-Xylenes	0.673	0.724	-7.6	86	0.00
69 T	o-Xylene	0.664	0.714	-7.5	88	0.00
70 T	Styrene	1.154	1.261	-9.3	87	0.00
71 P	Bromoform	0.295	0.324	-9.8	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
73 T	Isopropylbenzene	3.636	3.756	-3.3	86	0.00
74 T	N-amyl acetate	2.120	2.300	-8.5	88	0.00
75 P	1,1,2,2-Tetrachloroethane	1.355	1.431	-5.6	88	0.00
76 T	1,2,3-Trichloropropane	1.160	1.231	-6.1	89	0.00
77 T	Bromobenzene	0.902	0.944	-4.7	88	0.00
78 T	n-propylbenzene	4.152	4.317	-4.0	86	0.00
79 T	2-Chlorotoluene	2.562	2.597	-1.4	87	0.00
80 T	1,3,5-Trimethylbenzene	3.061	3.198	-4.5	87	0.00
81 T	trans-1,4-Dichloro-2-butene	0.453	0.435	4.0	77	0.00
82 T	4-Chlorotoluene	2.894	2.996	-3.5	87	0.00
83 T	tert-Butylbenzene	2.978	3.079	-3.4	87	0.00
84 T	1,2,4-Trimethylbenzene	3.073	3.239	-5.4	87	0.00
85 T	sec-Butylbenzene	3.531	3.661	-3.7	86	0.00
86 T	p-Isopropyltoluene	3.174	3.327	-4.8	86	0.00
87 T	1,3-Dichlorobenzene	1.609	1.667	-3.6	87	0.00
88 T	1,4-Dichlorobenzene	1.622	1.681	-3.6	87	0.00
89 T	n-Butylbenzene	2.869	2.926	-2.0	82	0.00

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051019\
Data File : VX009511.D
Acq On : 11 May 2019 03:04
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 11 04:41:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.541	0.562	-3.9	84	0.00
91 T	1,2-Dichlorobenzene	1.639	1.712	-4.5	88	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.304	0.322	-5.9	85	0.00
93 T	1,2,4-Trichlorobenzene	1.121	1.177	-5.0	85	0.00
94 T	Hexachlorobutadiene	0.572	0.586	-2.4	85	0.00
95 T	Naphthalene	3.646	3.907	-7.2	86	0.00
96 T	1,2,3-Trichlorobenzene	1.136	1.206	-6.2	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6