

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051019\
 Data File : VX009509.D
 Acq On : 11 May 2019 02:17
 Operator : JC/SP
 Sample : K2805-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: May 11 04:36:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	182751	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	290794	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	258384	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	112086	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	119788	48.02	ug/l	0.00
Spiked Amount			Recovery	=	96.04%	
35) Dibromofluoromethane	5.50	113	87389	47.80	ug/l	0.00
Spiked Amount			Recovery	=	95.60%	
50) Toluene-d8	8.71	98	364743	49.42	ug/l	0.00
Spiked Amount			Recovery	=	98.84%	
62) 4-Bromofluorobenzene	11.13	95	123170	45.93	ug/l	0.00
Spiked Amount			Recovery	=	91.86%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051019\
Data File : VX009509.D
Acq On : 11 May 2019 02:17
Operator : JC/SP
Sample : K2805-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: May 11 04:36:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.01 min

Lab File: VX009509.D

Acq: 11 May 2019 02:17

Instrument :

MSVOA_X

ClientSampleId :


STORAGE-BLANK-WATER-REF-4

Tot Ion:168 Resp: 182751

Ion Ratio Lower Upper

168 100

99 55.0 42.3 63.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.66



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 48.016 ug/l

RT: 6.06 min Scan# 805

Delta R.T. 0.00 min

Lab File: VX009509.D


Acq: 11 May 2019 02:17

Tgt Ion: 65 Resp: 119788

Ion Ratio Lower Upper

65 100

67 53.2 0.0 106.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.06

50000

40000


30000


20000

10000

0

Time--> 5.90 6.00 6.10 6.20









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. 0.00 min
 Lab File: VX009509.D
 Acq: 11 May 2019 02:17

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Tot Ion:114 Resp: 290794
 Ion Ratio Lower Upper
 114 100
 63 23.3 0.0 45.6
 88 16.9 0.0 33.2




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VX0
 Ion 87.90 (87.60 to 88.60): VX0






#35
 Dibromofluoromethane
 Concen: 47.805 ug/l
 RT: 5.50 min Scan# 713
 Delta R.T. 0.01 min
 Lab File: VX009509.D
 Acq: 11 May 2019 02:17

Tgt Ion:113 Resp: 87389
 Ion Ratio Lower Upper
 113 100
 111 103.2 82.8 124.2
 192 22.1 17.0 25.6



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 49.421 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. 0.00 min
Lab File: VX009509.D
Acq: 11 May 2019 02:17




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tot Ion: 98 Resp: 364743
Ion Ratio Lower Upper
98 100
100 63.9 51.2 76.8



#62
4-Bromofluorobenzene
Concen: 45.930 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. 0.00 min
Lab File: VX009509.D
Acq: 11 May 2019 02:17

Tgt Ion: 95 Resp: 123170
Ion Ratio Lower Upper
95 100
174 82.7 0.0 161.6
176 79.9 0.0 159.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. 0.00 min
Lab File: VX009509.D
Acq: 11 May 2019 02:17

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Tot Ion:117 Resp: 258384
Ion Ratio Lower Upper
117 100
82 60.3 48.8 73.2
119 31.8 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX009509.D
Acq: 11 May 2019 02:17

Tgt Ion:152 Resp: 112086
Ion Ratio Lower Upper
152 100
115 58.7 41.2 123.6
150 157.0 0.0 346.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

