


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX051019\
 Data File : VX009470.D
 Acq On : 10 May 2019 08:54
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\624X050119W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Thu May 02 06:46:57 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	6342	PASS
75	95	30	60	49.2	16003	PASS
95	95	100	100	100.0	32541	PASS
96	95	5	9	6.3	2041	PASS
173	174	0.00	2	0.4	102	PASS
174	95	50	100	78.4	25525	PASS
175	174	5	9	8.5	2169	PASS
176	174	95	101	95.9	24480	PASS
177	176	5	9	6.3	1543	PASS