

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051021\
 Data File : VX021886.D
 Acq On : 10 May 2021 19:33
 Operator : JC/MD
 Sample : VIBLK311
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK311

Quant Time: May 11 03:07:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR051021WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue May 11 02:59:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	83082	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	74552	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	34699	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	33757	4.73	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.60%
7) Chloroethane-d5	1.673	69	27616	4.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.60%
11) 1,1-Dichloroethene-d2	2.307	63	42805	3.61	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.20%
20) 2-Butanone-d5	4.465	46	77457	47.64	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.28%
24) Chloroform-d	5.062	84	55657	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.80%
26) 1,2-Dichloroethane-d4	5.964	65	29164	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.60%
32) Benzene-d6	5.983	84	117823	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.40%
36) 1,2-Dichloropropane-d6	7.312	67	35957	4.75	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.00%
41) Toluene-d8	8.653	98	106376	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.80%
43) trans-1,3-Dichloroprop...	8.952	79	15224	4.45	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.00%
46) 2-Hexanone-d5	9.385	63	69208	47.59	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.18%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	27099	4.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.60%
66) 1,2-Dichlorobenzene-d4	12.323	152	35486	4.88	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.60%
Target Compounds						Qvalue
42) Toluene	8.720	91	3105	0.11	ug/L	78
60) Isopropylbenzene	10.970	105	1591	0.05	ug/L	95
62) 1,3,5-Trimethylbenzene	11.457	105	1552	0.06	ug/L	95
63) 1,2,4-Trimethylbenzene	11.750	105	1679	0.07	ug/L	94
64) 1,3-Dichlorobenzene	11.969	146	1256	0.10	ug/L	93
65) 1,4-Dichlorobenzene	12.049	146	1224	0.10	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051021\
Data File : VX021886.D
Acq On : 10 May 2021 19:33
Operator : JC/MD
Sample : VIBLK311
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK311

Quant Time: May 11 03:07:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR051021WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue May 11 02:59:37 2021
Response via : Initial Calibration

