

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
 Data File : VX028584.D
 Acq On : 10 May 2022 10:47
 Operator : JC/MD
 Sample : VX0510WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0510WBL01

Quant Time: May 11 06:56:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	450436	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	754165	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	712285	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	343336	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	299590	51.010	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.020%	
35) Dibromofluoromethane	5.385	113	281154	53.159	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 106.320%	
50) Toluene-d8	8.647	98	1012636	49.850	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 99.700%	
62) 4-Bromofluorobenzene	11.079	95	369419	47.562	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 95.120%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
Data File : VX028584.D
Acq On : 10 May 2022 10:47
Operator : JC/MD
Sample : VX0510WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0510WBL01

Quant Time: May 11 06:56:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

