

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
 Data File : VX028590.D
 Acq On : 10 May 2022 13:16
 Operator : JC/MD
 Sample : PB144702ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144702ZHE#03

Quant Time: May 11 06:59:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	418545	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	706079	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	695104	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	334915	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	285600	52.333	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 104.660%			
35) Dibromofluoromethane	5.391	113	266448	53.810	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 107.620%			
50) Toluene-d8	8.653	98	977680	51.407	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 102.820%			
62) 4-Bromofluorobenzene	11.079	95	362029	49.785	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 99.580%			
Target Compounds					Qvalue	
16) Acetone	2.380	43	40882	23.260	ug/l	95
20) Methylene Chloride	2.794	84	36631	9.046	ug/l	99
43) Isopropyl Acetate	6.342	43	29730	2.787	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
Data File : VX028590.D
Acq On : 10 May 2022 13:16
Operator : JC/MD
Sample : PB144702ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144702ZHE#03

Quant Time: May 11 06:59:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

