

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
 Data File : VX028591.D
 Acq On : 10 May 2022 13:39
 Operator : JC/MD
 Sample : PB144702ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144702ZHE#04

Quant Time: May 11 06:59:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	413193	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	704115	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	693809	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	341434	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	289144	53.669	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.340%	
35) Dibromofluoromethane	5.391	113	269898	54.658	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 109.320%	
50) Toluene-d8	8.653	98	985496	51.963	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.920%	
62) 4-Bromofluorobenzene	11.079	95	370126	51.040	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 102.080%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	40346	23.252	ug/l	100
20) Methylene Chloride	2.794	84	36871	9.224	ug/l	97
43) Isopropyl Acetate	6.348	43	30294	2.848	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
Data File : VX028591.D
Acq On : 10 May 2022 13:39
Operator : JC/MD
Sample : PB144702ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144702ZHE#04

Quant Time: May 11 06:59:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

