

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
 Data File : VX028592.D
 Acq On : 10 May 2022 14:02
 Operator : JC/MD
 Sample : PB144702ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144702ZHE#05

Quant Time: May 11 06:59:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	406757	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	685297	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	673537	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	317454	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	275339	51.915	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 103.840%			
35) Dibromofluoromethane	5.391	113	259509	53.998	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 108.000%			
50) Toluene-d8	8.653	98	945458	51.220	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 102.440%			
62) 4-Bromofluorobenzene	11.079	95	349208	49.478	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 98.960%			
Target Compounds						
16) Acetone	2.380	43	27912	15.924	ug/l	99
20) Methylene Chloride	2.788	84	16582	4.214	ug/l	99
43) Isopropyl Acetate	6.348	43	29201	2.821	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
Data File : VX028592.D
Acq On : 10 May 2022 14:02
Operator : JC/MD
Sample : PB144702ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144702ZHE#05

Quant Time: May 11 06:59:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

