

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
 Data File : VX028594.D
 Acq On : 10 May 2022 14:49
 Operator : JC/MD
 Sample : PB144702ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144702ZHE#07

Quant Time: May 11 07:00:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	413334	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	707587	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	698815	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	336284	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	284370	52.765	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 105.520%	
35) Dibromofluoromethane	5.385	113	264231	53.248	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 106.500%	
50) Toluene-d8	8.653	98	977953	51.312	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.620%	
62) 4-Bromofluorobenzene	11.079	95	363470	49.877	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 99.760%	
Target Compounds						Qvalue
16) Acetone	2.380	43	28367	15.926	ug/l	97
20) Methylene Chloride	2.788	84	16839	4.211	ug/l	93
43) Isopropyl Acetate	6.342	43	29366	2.747	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
Data File : VX028594.D
Acq On : 10 May 2022 14:49
Operator : JC/MD
Sample : PB144702ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144702ZHE#07

Quant Time: May 11 07:00:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

