

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
 Data File : VX028595.D
 Acq On : 10 May 2022 15:12
 Operator : JC/MD
 Sample : PB144702ZHE#08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144702ZHE#08

Quant Time: May 11 07:00:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	400038	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	690081	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	678711	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	331882	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	280828	53.839	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.680%	
35) Dibromofluoromethane	5.391	113	265632	54.888	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 109.780%	
50) Toluene-d8	8.653	98	971993	52.293	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.580%	
62) 4-Bromofluorobenzene	11.079	95	361646	50.885	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.780%	
Target Compounds						Qvalue
16) Acetone	2.380	43	27936	16.230	ug/l	98
20) Methylene Chloride	2.788	84	17075	4.412	ug/l	97
43) Isopropyl Acetate	6.342	43	29698	2.849	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
Data File : VX028595.D
Acq On : 10 May 2022 15:12
Operator : JC/MD
Sample : PB144702ZHE#08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144702ZHE#08

Quant Time: May 11 07:00:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

