

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
 Data File : VX028598.D
 Acq On : 10 May 2022 16:45
 Operator : JC/MD
 Sample : PB144702ZHE#12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144702ZHE#12

Quant Time: May 11 07:01:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	403007	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	697473	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	702096	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	345216	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	281248	53.523	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.040%	
35) Dibromofluoromethane	5.391	113	259742	53.103	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 106.200%	
50) Toluene-d8	8.653	98	972803	51.782	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.560%	
62) 4-Bromofluorobenzene	11.079	95	366182	50.977	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.960%	
Target Compounds						
						Qvalue
16) Acetone	2.373	43	43433	25.810	ug/l	95
20) Methylene Chloride	2.788	84	36563	9.378	ug/l	97
43) Isopropyl Acetate	6.342	43	29984	2.846	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051022\
Data File : VX028598.D
Acq On : 10 May 2022 16:45
Operator : JC/MD
Sample : PB144702ZHE#12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144702ZHE#12

Quant Time: May 11 07:01:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

